



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/752,435

01/06/2004

Lee Bolduc

9494.18510

3762

26308 7590 12/08/2008
RYAN KROMHOLZ & MANION, S.C.
POST OFFICE BOX 26618
MILWAUKEE, WI 53226

EXAMINER

SEVERSON, RYAN J

ART UNIT

PAPER NUMBER

3731

MAIL DATE

DELIVERY MODE

12/08/2008

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/752,435	Applicant(s) BOLDUC ET AL.	
	Examiner Ryan Severson	Art Unit 3731	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 19 September 2008.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 23,28 and 31-34 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 23,28 and 31-34 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 09 January 2008 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date <u>9/19/2008</u> . | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 103

1. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

2. **Claims 23 and 28 are rejected under 35 U.S.C. 103(a) as being unpatentable over Parodi et al. (WO 00/16701) in view of Taheri et al. (5,591,195), Pinchuk (5,855,598) and Layne et al. (6,398,803).** Parodi reference discloses a fastener applier (50) is used to secure a prosthesis (100) using helical fasteners (80, see figure 6). Multiple fasteners are deployed (see figure 3) to secure the prosthesis at an aneurysm in an aorta (see figure 2 and page 9, lines 18-22).

3. However, Parodi does not disclose the prosthetic has a scaffold that supports the prosthetic material. Attention is drawn to Taheri reference, which teaches a prosthetic material can be supported by a scaffold (see figure 1) to provide greater resistance to collapse of the vessel if the vessel becomes too weakened. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to include the scaffold of Taheri reference with the prosthetic material of Parodi reference to provide greater resistance to collapse of the vessel if the vessel becomes too weakened.

4. Regarding claim 23, the combination of Parodi and Taheri references does not disclose a first and second prosthesis telescopically positioned relative to one another.

Art Unit: 3731

Attention is drawn to Pinchuk reference, which teaches it is known in the art to use a telescopically fitted multiple-component prosthesis (see figures 18-21) to allow the prosthesis to support the vessel from the neck region adjacent the renal arteries to the iliac arteries. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to make the prosthesis of the combination of Parodi and Taheri references a multiple-component prosthesis to provide for support of the iliac arteries as well as the aneurysm in the aorta.

5. Further regarding claim 23, the combination of Parodi, Taheri, and Pinchuk references does not disclose the graft material has a greater density at the ends than in the intermediate region. Attention is drawn to Layne reference, which teaches a graft may have more dense ends (because the intermediate region has openings 44) to increase the flexibility of the prosthesis in the intermediate region. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to make the intermediate region of the graft of the combination of Parodi, Taheri, and Pinchuk references less dense in the manner taught by Layne reference to increase the flexibility of the prosthesis in the intermediate region.

6. **Claims 32 and 33 are rejected under 35 U.S.C. 103(a) as being unpatentable over Parodi et al. (WO 00/16701) in view of Taheri et al. (5,591,195) and Pinchuk (5,855,598).** The combination of Parodi, Taheri, and Pinchuk reference discloses the limitations of the claims as set forth in paragraphs 2-4 above. Further, Taheri shows the scaffold structure is less dense (fewer struts per area) in the intermediate portion of the prosthesis (see figure 1).

Art Unit: 3731

7. **Claim 31 is rejected under 35 U.S.C. 103(a) as being unpatentable over Parodi et al. (WO 00/16701) in view of Taheri et al. (5,591,195), Pinchuk (5,855,598) and Layne et al. (6,398,803) as applied to claim 23 above, and further in view of Goicoechea et al. (5,609,627).** The combination of Parodi, Taheri, Pinchuk, and Layne references does not disclose the first trunk has fluoroscopic markers. Attention is drawn to Goicoechea reference, which teaches markers (see column 10, lines 53-63) on a stent to ensure proper placement of the stent in the vessel. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to have included markers on the stent of the combination of Parodi, Taheri, Pinchuk, and Layne references in the manner taught by Goicoechea reference to ensure proper placement of the stent in the vessel.

8. **Claim 34 rejected under 35 U.S.C. 103(a) as being unpatentable over Parodi et al. (WO 00/16701) in view of Taheri et al. (5,591,195) and Pinchuk (5,855,598) as applied to claim 32 above, and further in view of Goicoechea et al. (5,609,627).**

The combination of Parodi, Taheri, and Pinchuk references does not disclose the first trunk has fluoroscopic markers. Attention is drawn to Goicoechea reference, which teaches markers (see column 10, lines 53-63) on a stent to ensure proper placement of the stent in the vessel. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to have included markers on the stent of the combination of Parodi, Taheri, and Pinchuk references in the manner taught by Goicoechea reference to ensure proper placement of the stent in the vessel.

Response to Arguments

9. Applicant's arguments with respect to claims 23 and 32 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

10. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

11. A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

12. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Ryan Severson whose telephone number is (571)272-3142. The examiner can normally be reached on Monday - Friday 8:30-5:00.

13. If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Todd Manahan can be reached on (571) 272-4713. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3731

14. Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/R. S./
Examiner, Art Unit 3731

/Todd E Manahan/
Supervisory Patent Examiner, Art Unit 3731